Research Article

Rus

Methodical approaches to bioassay of substances containing unstable functional groups

Alexander L. Khokhlov¹, Ilya I. Yaichkov^{1,2}, Yuriy A. Dzhurko³, Leonid N. Shitov^{1,3}, Anastasia A. Shitova³

- 1 Yaroslavl State Medical University, 5 Revolyutsionnaya St., Yaroslavl 150000 Russia
- 2 Yaroslavl State Pedagogical University named after K.D. Ushinsky, 108/1 Respublikanskaya St., Yaroslavl 150000 Russia
- 3 Quinta-Analytica Yaroslavl LLC, 52g Leningradskaya St., Yaroslavl 150045 Russia

Corresponding author: Ilya I. Yaichkov (ilya_1993_08@mail.ru)

Academic editor: Mikhail Pokrovskii • Received 13 February 2018 • Accepted 21 March 2018 • Published 28 March 2018

Citation: Khokhlov AL, Yaichkov II, Dzhurko YA, Shitov LN, Shitova AA (2018) Methodical approaches to bioassay of substances containing unstable functional groups. Research Result: Pharmacology and Clinical Pharmacology 4(1): 33–42. https://doi.org/10.3897/rrpharmacology.4.25253

Abstract

Introduction: This article describes the method development approaches for bioassay of substances containing unstable functional groups and forming unstable metabolites using the example of mycophenolic acid, methyldopa and mebeverine metabolites.

Materials and Methods: The concentration of mycophenolic acid, which contains one phenolic hydroxyl and forms glucuronides during metabolism, was measured in plasma using HPLC-MS/MS, HPLC-MS and GC-MS. The determination of methyldopa, containing two phenolic hydroxyls, in stabilised plasma was performed by HPLC-MS/MS in the range of 0.02-3.00 μg/ml. Desmethyl mebeverine acid, which contains one phenolic hydroxyl and is metabolised by forming phenolic glucuronide, was assayed simultaneously with mebeverine acid in the range of 10-2000 ng/ml.

Results and Discussion: The selection of storage conditions of the samples containing unstable substances should begin with selecting an anticoagulant based on the study of its short-term stability and freeze/thaw stability. If an unacceptable result is obtained, a combination of the anticoagulant and a stabiliser solution, as well as a concentration of this solution and its volume ratio to the biological fluid should be titrated. After which, this method should be validated by using the selected anticoagulant or the combination of the anticoagulant and stabiliser solution.

Conclusion: The application of this approach to developing a bioanalytical method for determination of unstable compounds makes it possible to avoid obtaining false assay results.

Keywords

bioanalytical studies, instability, mycophenolic acid, methyldopa, desmethyl mebeverine acid

Introduction

Stability of the analytes in biological matrices is an important component, which guarantees obtaining reliable results of bioanalytical studies. Thus, a wrong conclusion about the bioequivalence (BE) of the generic and original drug can lead to the registration of the drug which does

not meet the requirements of effectiveness and safety. Errors of therapeutic drug monitoring may result in a faulty dosage (Khokhlov 2017), which puts patients' health at high risk.

Oxidation and hydrolysis are the main causes for the decomposition of molecules of drugs and their metabolites in biological fluids (Khokhlov 2017, Yaichkov et al.

2017, Dell 2004, Khokhlov et al. 2017a, Hilhorst et al. 2017, Li et al. 2013, Li et al. 2011). Stabilisation of biological samples may be achieved by changing the storage conditions, selecting an anticoagulant and adding stabilisers depending on the structure of the analyte: antioxidants, buffer solutions, acid solutions and bases for pH correction, enzyme inhibitors (Yaichkov et al. 2017, Dell 2004, Khokhlov et al. 2017a, Hilhorst et al. 2017, Li et al. 2013, Li et al. 2011). Substances which contain phenolic hydroxyls are the main examples of readily oxidisable compounds (Khokhlov et al. 2017a, Hilhorst et al. 2017). Back conversion of glucurone conjugates into the parent drugs is the most common example of the hydrolysis of compounds in biological fluids (Hilhorst et al. 2017, Li et al. 2011). Thus, over the last 5 years, results have been published for over 30 bioanalytical studies of drugs which metabolise by forming glucuronides (Yaichkov et al. 2017, Li et al. 2011, Brktas et al. 2016, Veeragoni et al. 2016, Shah et al. 2016, Sengupta et al. 2017, Patel et al. 2017). The least stable are O-acylglucuronides (Yaichkov et al. 2017, Dell 2004, Hilhorst et al. 2017, Li et al. 2013, Li et al. 2011); however, phenolic glucuronides may also undergo hydrolytic decomposition (Li et al. 2011, Ye et al. 2017).

Therefore, mycophenolic acid (MPA) and desmethyl mebeverine acid (DMA) which both contain one phenolic hydroxyl and methyldopa (MD) which contains two phenolic hydroxyls (Fig. 1) were selected to develop approaches to bioassay of substances containing unstable functional groups.

Besides, MPA and DMA are conjugated to glucuronides during metabolism (Khokhlov 2017, Kristinsson et al. 1994, Khokhlov et al. 2016), which results in MPA forming acylglucuronide (AcMPAG) and phenolic glucuronide (MPAG) with DMA forming only phenolic glucuronide (DMAG). A literature review showed that there was no need to add antioxidant solutions to the samples of biological fluids (Khokhlov et al. 2016, Benoit-Biancamano et al. 2007, Brandhorst et al. 2016, Heinig et al. 2010, Oliveira et al. 2002, Ryna et al. 2001, Vlase et al. 2013, Valizadeh et al. 2010, Elliott and Burgess 2006). The results of studying back-conversion of MPAG and AcMPAG during the storage process differ: some authors emphasise the necessity to use buffer solutions to slow hydrolysis of these metabolites (Benoit-Biancamano et al. 2007, Brandhorst et al. 2016), whereas others state that degradation of MPA conjugates was not significant and adding stabilisers was not necessary (Khokhlov et al. 2016, Heinig et al. 2010). Therefore, studying this process is of current concern.

Materials and Methods

Sample preparation for HPLC-MS/MS-determination of MPA was performed using deproteinisation: 450 μ l of the deuterated internal standard (IS) methanol solution of mycophenolic acid-D3 was added to 50 μ l of plasma. The mixture was vortexed and centrifuged at 2500 rpm for 10 min. A Kinetex C18 column (30 * 4.6 mm, 2.6 μ m) with

Figure 1. Graphic formulae of mycophenolic acid (A), methyldopa (B) and desmethyl mebeverine acid (C).

a Phenomenex Security guard (C18, 4 * 3 mm) and a mobile phase based on acetonitrile and water in the gradient elution mode were applied for the analysis. Mass spectrometry detection was carried out in negative ion MRM mode by the following MRM-transitions: MPA – 319 \rightarrow 191+205 m/z; MPA-D₃ – 322 \rightarrow 191 + 205 m/z (Khokhlov et al. 2016).

Protein precipitation was also used for preparing the samples for HPLC-MS-determination of MPA: 50 µl of plasma was vortexed with 200 µl of methanol. The mixture was centrifuged at 10000 rpm for 5 min. Separation was conducted on a Zorbax Eclipse Plus C18 column (100 * 4.6 mm, 3.5 µm) using a mobile phase based on acetonitrile, water and a 0.1% solution of formic acid

(50:45:5, v/v/v) in the isocratic elution mode. Negative ions of MPA were detected in the SIM-mode by molecular ion -319 m/z (Khokhlov et al. 2017a).

Liquid-liquid extraction by methylene chloride after correcting the pH value to 2.0 by phosphate buffer solution was used for preparing plasma samples for GC-MS-determination of MPA. After separation and evaporation of the extract, derivatisation of the analyte by a mixture of N,O-bis (trimethylsilyl)-trifluoroacetamide and trimethylchlorosilane was performed. Chromatographic separation was done using Mega 5-MS column (25 m * 0.20 mm, 0.33 μm) under the following temperature programme: the initial temperature was 100°C for 3 min, heating was done at a rate of 25°C/min up to 300εC, the final temperature

was 300ϵ C for 8 min (the total run time - 19 min). Mass spectrometric detection of di-TMS-derivative of MPA was performed in the SIM-mode – 449 m/z.

The method of HPLC-MS/MS was used to measure the concentration of DMA and a minor mebeverine metabolite – mebeverine acid (MA) in plasma. The samples were prepared by protein precipitation. A methanol solution (400 µl) of deuterated standards of mebeverine acid-D_e $(MA-D_{\varepsilon})$ and desmethyl mebeverine acid-D_{\varepsilon} $(DMA-D_{\varepsilon})$ was added to an aliquot of plasma of 100 μl. The mixture was centrifuged at 3500 rpm for 10 min after vortexing. Chromatographic separation was performed on two columns Luna C8 Mercury (20*4.0 mm, 5 μm) and Luna 5u C8 (150*4.6 mm, 5 μm) using a mobile phase based on acetonitrile, methanol and formate buffer solution in the gradient elution mode. Mass spectrometry detection was carried out in positive ion registration mode by the following MRM-transitions: MA – 280→121 m/z, DMA $-266 \rightarrow 107 \text{ m/z}, \text{MA-D}_{5} - 285 \rightarrow 121 \text{ m/z} \text{ and DMA-D}_{5} 271 \rightarrow 107 \text{ m/z}$ (Khokhlov et al. 2017a).

The samples containing methyldopa for HPLC-MS/MS determination of its concentration were prepared by protein precipitation. First, 400 μ l of deuterated internal standard of methyldopa-D₃ (MD-D₃) in methanol was added to 100 μ l of plasma; then the mixture was vortexed and centrifuged at 3500 rpm and temperature of +4°C for 10 min. Chromatographic separation of the sample components was done using two columns Luna Phenyl-Hexyl (50*3.0 mm, 5 μ m) and Synergi Fusion RP 80E (150*3.0 mm, 4 μ m) in the isocratic elution mode. Mass spectrometry detection was carried out in positive ion registration mode by the following MRM-transitions: for MD – 212 \rightarrow 139 m/z; for MD-D₃ (IS) 215 \rightarrow 169 m/z (Khokhlov et al. 2017c).

Results and Discussion

At the beginning of developing the methods for determining MPA using HPLC-MS and HPLC-MS/MS, back-conversion of MPAG in the ion source was investigated. The selected chromatographic parameters for both methods allow separating the analyte and its main metabolite (Figs. 2 and 3); that is the reason why its fragmentation in the ionisation process does not affect the accuracy of measuring MPA. This experiment was not carried out when using the GC-MS method, because MPAG is not recovered from the plasma under the selected extraction conditions.

A preliminary evaluation of the stability of mycophenolic acid was made by applying the HPLC-MS/MS method in plasma samples at a concentration of 25.0 μg/ml using K₃EDTA and lithium heparinate as anticoagulants. The results obtained after 24 h of keeping the model mixtures at room temperature and 3 freeze/thaw cycles meet the acceptance criteria: mean concentration values of MPA at K₃EDTA were 99.8% and 100.6% of the initial concentration, respectively and at lithium heparinate-plasma they were 101.1% and 97.2% of the initial concentration, respectively. Thus, mycophenolic acid containing one phenolic hydroxyl is resistant to oxidation in blood plasma.

Back-conversion of MPAG during the storage process was also studied using these anticoagulants on the samples at a concentration of this metabolite of 100 µg/ml (Khokhlov et al. 2016). If the lower limit of quantification (LLOQ) of the method is 0.05 µg/ml, MPAG hydrolysis exceeds the maximum permissible level (20% of the MPA chromatographic peak area of the LLOQ sample) after 6 hours of storage at room temperature using K,ED-TA, which is significantly higher than with using lithium heparinate (Table 1). If LLOQ of the method is $0.5 \mu g$ / ml, back-conversion of MPAG is at the acceptable level when applying both anticoagulants. The LLOQ level of 0.5 µg/ml is sufficient for conducting of BE studies of mycophenolic acid formulations (Khokhlov et al. 2016). However, when using lithium heparinate, the hydrolysis degree of the metabolite was approximately three times higher. That is the reason why K,EDTA was selected for further studies (Table 1).

The absence of influence of back-conversion of AcM-PAG and MPAG on the accuracy of mycophenolic acid determination was also proved by re-analysis of the plasma samples obtained from rats for cross-validation. The difference between the initial and final results of measurements ranged from -3.43% to 9.49%, which does not exceed the maximum acceptable value of 20% (European Medicines Agency 2010, Mironov 2014, Council of the Eurasian Economic Commission 2016).

When developing a method for determining mebeverine metabolites in plasma, fragmentation of DMAG in the ionisation process was also studied. There was no decomposition of this metabolite in the ion source (Fig. 4).

The results of testing the stability of desmethyl mebeverine acid, containing one phenolic hydroxyl, in plasma after 24 hours by maintaining samples at room temperature and 3 freeze/thaw cycles, met the acceptance criteria. Back-conversion of DMAG in the samples with a concentration of 2000 ng/ml kept for 24 hours at room tem-

Table 1. Study of Back-Conversion of Phenolic Glucuronide of Mycophenolic Acid During Storage.

			% of the chrom. peak ar	ea of LLOQ samp	ole
		LLOQ of the	e method 0,5 μg/ml	LLOQ of the	method 0,05 μg/ml
		K ₃ EDTA	Lithium heparinate	K ₃ EDTA	Lithium heparinate
Storage time, h	2 h	-	-	6.0	23.9
(room temperature)	6 h	-	-	17.0	75.2
	24 h	2.58	7.44	45.2	154.3




Figure 2. MRM-chromatogram of the mixture of MPA and MPAG when using HPLC-MS/MS method.




Figure 3. Chromatograms of MPA (A) and MPAG (B) solutions when using HPLC-MS method.




Figure 4. Chromatograms of DMAG (A) and DMA (B) solutions.

perature was completely absent. Consequently, addition of the stabiliser solution to the K₃EDTA plasma in this case is not required.

A preliminary study of the short-term stability and freeze/thaw stability of methyldopa, containing two phenolic hydroxyls, was carried out on plasma model mixtures at a concentration of 2.40 µg/ml. K₂EDTA and lithium heparinate were used as anticoagulants. The obtained results did not meet the acceptance criteria (Table 2). Therefore, it was necessary to add an antioxidant solution to the plasma to prevent oxidation of the analyte. The usage of solutions of ascorbic acid at a concentration of 5 and 10% and a solution containing a mixture of ascorbic acid, sodium sulphite and sodium bicarbonate at concentrations of 5%, 0.2% and 2.4%, respectively, used with K₂EDTA in the ratio of 1:5 (antioxidant solution/plasma), made it possible to prevent degradation of the substance (Table 2). The mixture of ascorbic acid, sodium sulphite and sodium bicarbonate was selected for further studies, because the sensitivity of the method was highest when using this mixture (Khokhlov et al. 2017a, Khokhlov et al. 2017c).

Thus, addition of an antioxidant stabiliser was necessary only when studying methyldopa, which contains two phenolic hydroxyls. The DMA molecule remained stable in plasma, despite the absence in the benzene ring of electron-withdrawing groups, which reduce the electron density and, thus, oxidation capacity. Back-conversion of the phenolic glucuronide of DMA was also completely absent. Hydrolysis of MPA conjugates, when using K₃EDTA as an anticoagulant, was at an acceptable level, which made it possible to avoid using buffer solutions to correct the pH of the medium.

Validation of the developed methods

The developed methods were validated in accordance with the requirements of EMA Guidelines (European Medicines Agency 2010), The Guidelines on the Eva-

		Short-term stability (24 h at room temperature), n=2	Freeze/ thaw stability n=2	Short-term stability (24 h at room temperature), n=2	Freeze/ thaw stability n=2
	Concentration of			concentration	11 2
Stabiliser	stabiliser, %	K,E	DTA	Lithium h	eparinate
Without stabiliser	-	30.29	80.04	47.16	93.18
Ascorbic acid	5	88.44	102.30	80.91	99.69
	10	94.91	96.17	78.76	93.21
Mixture of ascorbic acid,	5% ascorbic acid, 0.2%				
sodium bicarbonate,	sodium sulphite, 2.4%	89.31	98.24	72.48	95.07
sodium sulphite	sodium bicarbonate				
Sodium thiosulphate	5	45.65	101.86	79.41	93.91
	10	31.32	99.81	83.46	97.78
Sodium metabisulphite	5	66.84	102.68	75.79	90.70
	10	75.99	97.25	73.22	88.70

luation of Medicinal Products by The Scientific Centre for Expert Evaluation of Medicinal Products of the Ministry of Healthcare of the Russian Federation (Vol. 1) (Mironov 2014) and Decision of the Council of the Eurasian Economic Commission № 85 "On Approval of the Rules for Conducting Bioequivalence Studies on Medicines in the Eurasian Economic Union" (Council of the Eurasian Economic Commission 2016). The results of validation tests meet all the acceptance criteria (Table 3).

Thus, the selected anticoagulants (for MPA and DMA) and the combination of an anticoagulant with a stabiliser solution (for MD) provide stability of the analytes in plasma samples for 24 hours at room temperature, 3 freeze/ thaw cycles and under freezing conditions (Table 4). The usage of a storage temperature not higher than -80°C made it possible to increase the storage period of samples containing methyldopa by up to 3 months.

Application of the developed methods

The HPLC-MS/MS method of MPA determination was applied for conducting a bioequivalence study of coated tablets of mycophenolate sodium (Khokhlov et al. 2016). The HPLC-MS and GC-MS-methods of quantification of mycophenolic acid were tested by cross validation in rat plasma samples. The method developed for measuring MD concentrations in plasma was used in an open randomised cross-over study of bioequivalence of methyldopa tablet formulations (Khokhlov et al. 2017a). The method of determining MA and DMA in plasma was applied to conduct a pharmacokinetics study of mebeverine sustained release capsules (Khokhlov et al. 2017b).

The statistical analysis was performed using StatSoft STATISTICA v.10, R package, Bear module (Lee, Hsinya and Lee, Yung-jin, bear: Data Analysis Tool for Average Bioequivalence and Bioavailability) and Microsoft Excel 2007.

Results of cross validation of methods for determining mycophenolic acid

The study was carried out on 10 white mongrel male rats weighing 250±10 g in accordance with the principles of Good Laboratory Practice (Ministry of Healthcare of the Russian Federation 2016). The aqueous solution of mycophenolate sodium was administered to rats orally in the dosage of 33.0 mg/kg. Blood sampling was performed 2 h after administration, as it is the time of MPAG and AcM-PAG peak concentrations in plasma (Liu et al. 2017). Part of the plasma obtained after centrifusion was subjected to immediate sample preparation. The HPLC-MS/MS-method was chosen as a reference method, because to calculate the concentrations of MPA, a stable isotope-labelled internal standard MPA-D3 was used which is the most preferable in accordance with the Eurasian Economic Union Guidelines for conducting BE studies (Council of the Eurasian Economic Commission 2016).

The relative concentrations of MPA obtained by the HPLC-MS-method were within the range of 95.18-106.93% compared to the HPLC-MS/MS results; concentrations obtained by the GC-MS-method were within the range of 94.27-110.26% (Table 5). It meets the acceptance criteria (European Medicines Agency 2010, Mironov 2014, Council of the Eurasian Economic Commission 2016): relative error does not exceed \pm 20% for over 67% of samples. Thus, HPLC-MS and GC-MS methods developed for MPA determination provide accurate results without using a stable isotope-labelled internal standard for calculating the concentration.

Results of bioequivalence study of methyldopa tablet formulations

The study of comparative pharmacokinetics of methyldopa formulations was conducted on 24 healthy volunteers. The test drug was Methyldopa tablets (250 mg) (R-Pharm,

Table 3. Validation Results of Methods Developed.

		Mycophenolic acid		Mebeverine meta	Mebeverine metabolites (HPLC-MS/MS)	Methyldopa (HPLC-
Parameter	HPLC-MS/MS	HPLC-MS	GC-MS	Mebeverine acid	Mebeverine acid Desmetyl mebeverine acid	MS/MS)
0-144	Interference in the rete	ntion times of analytes did	not exceed 20% of the L	LOQ level, in the retention	Interference in the retention times of analytes did not exceed 20% of the LLOQ level, in the retention times of internal standards (except for HPLC-MS and	ept for HPLC-MS and
Selectivity	GC-MS methods of de	GC-MS methods of determining MPA) did not exceed 5% of the average peak area.	ceed 5% of the average p	eak area.		
LLOQ	0.5 µg/ml	0.05 µg/ml	0.05 µg/ml	10 ng/ml	10 ng/ml	0.02 µg/ml
Calibration curve	0.5-30.0 µg/ml	0.05-30.00 µg/ml	0.05-30.00 µg/ml	10-2000 ng/ml	10-2000 ng/ml	0.02-3.00 µg/ml
Accuracy (relative error)) -4.03 ¤ +12.85%.	-14.02% ч +11.82%	-14.35% q +6.82%	-11.83% ч +13.86%	-1.73% u +13.77%	-5.28% q +7.42%
Precision (CV, %)	1.72% ч 8.52%	0.36% ч 9.82%	1.95% ч 5.58%	1.40% ч 8.24%	1.40%ч 8.58%	0.69% ч 3.97%
Matrix effects LQC	7.46%	2.96% (MF)	2.28% (MF)	3.66%	1.59%	0.38%
(CV NMF) HQC	6.55%	5.60% (MF)	2.80% (MF)	5.47%	2.82%	1.45%
Dilution integrity (relative	ve -6.47% q +5.71%	-9.46% 4 +4.22%	-0.84% q +10.99%	-2.04% 4+1.88%	+0.64% 4 +4.13%	-1.94% 4 +3.27%
error)						

Note: LLOQ - lower limit of quantification, LQC- low QC samples; HQC - high QC samples; NMF - normalised matrix factor; MF- matrix factor

Table 4. Stability Study of Analytes in Blood Plasma.

						Mebeverine metabolites	metabolites			
Parameter			Mycophenolic acid	nolic acid	MA	A	DMA	[A	Methyldopa	dopa
Short-term stability	ility	ТОС	99.18	18	107.72	.72	106.33	33	95.50	20
(24 h), % initial conc.	l conc.	НОС	99.46	46	97.81	81	98.54	54	89.58	82
Freeze/ thaw stability	ability	LQC	105.10	.10	111	111.83	105.44	44	104.43	43
(3 cycles), % initial conc.	nitial conc.	НОС	103.55	.55	102.45	.45	102.79	62	106.95	95
Autosampler stability,	ability,	ТОС	107.72	.72	109.06	90:	103.44	44	101.91	91
% initial conc.		НОС	108.07	.07	97.44	44	94.96	96	101.28	28
Long-term	not exceeding	LQC	100.38	103.24	111.56	109.33	108.61	105.83	90.74	66.92
stability, %	-20°C		(1 month)	(4 months)	(1 month)	(4 months)	(1 month)	(4 months)	(1 month)	(3 months)
initial conc.		НОС	94.79	100.86	105.45	103.56	110.33	110.06	86.20	59.57
			(1 month)	(4 months)	(1 month)	(4 months)	(1 month)	(4 months)	(1 month)	(3 months)
	not exceeding	Γ OC	ı	ı	I	ı	I	I	100.60 (1	100.76 (3
	−80°C								month)	months)
		НОС	ı	ı	ı	ı	1	ı	99.12	94.08
									(1 month)	(3 months)

Russia, batch number 20614, expiry date: 07.2016) and a reference preparation was Dopegit (Egis, Hungary, batch number D358N1014, expiry date: 10.2019). Blood samples was taken pre-dose and post-dose at 5 min, 15 min, 30 min, 45 min, 1 h, 1 h 15 min, 1 h 30 min, 1 h 45 min, 2 h, 2 h 15 min, 2 h 30 min, 3 h, 4 h, 6 h, 9 h, 12 h, 18 h and 24 h after drug administration in vacuum tubes with K₃EDTA. Then the blood samples were immediately centrifuged at a speed of 3000 rpm and a temperature of +4°C for 5 minutes. A stabiliser solution was added to the obtained plasma. A total of 864 samples were analysed using the developed method.

The values of the pharmacokinetic parameters of the drugs under study are not statistically significantly different (Table 6) and their average pharmacokinetic profiles practically coincide (Fig. 5). The calculated values of the 90%-confidence intervals of the parameters $C_{\rm max}$, $AUC_{\rm 0-t}$, $C_{\rm max}/AUC_{\rm 0-t}$ are within the acceptable limits of 80.0-125.0% (Table 7). The coefficient of intra-subject variability of $C_{\rm max}$ was 33.10%, which conformed to the level of a highly variable drug. The results of multivariate analysis of variance (ANOVA) show that the main factor contributing to the variability of this parameter was the

subjects themselves (p = 0.027). Incurred samples re-analysis (of 96 samples - 11.1% of the total amount) showed 100% reproducibility of the analytical method.

Thus, Methyldopa (by R-Pharm, Russia) and Dopegit (by Egis, Hungary) are bioequivalent. The measures developed for stabilisation of the analyte ensure validity of the results obtained.

Results of the pharmacokinetic study of mebeverine sustained release capsules

The cohort for assessing pharmacokinetic parameters of MA and DMA after administration of sustained release capsules of Duspatalin (200 mg) (by Abbott Healthcare SAS, France, batch number 10215, expiry date: 04.17) included 24 subjects aged 18-45 years. Blood samples were taken pre-dose and post-dose at 0.5 h, 1 h, 1.5 h, 2 h, 2.5 h, 3 h, 3.5 h, 4 h, 5 h, 6 h, 8 h, 10 h, 12 h and 24 h after drug administration in vacuum tubes with K₃EDTA. Then the plasma was immediately separated by centrifusion and frozen. A total of 360 samples were analysed.

Thus, the obtained data about the pharmacokinetic parameters of mebeverine metabolites after taking sustained release capsules can be used to develop a bioequivalence study design of studies (Table 8, Fig. 6)

Table 5. Cross Validation of Methods for Determining Mycophenolic Acid in Plasma.

	Determinations	Determinations by l	HPLC-MS method	Determinations by	GC-MS-method
	by HPLC-MS/MS				
№ of sample	method	Abs. μg/ml	Rel. %	Abs. μg/ml	Rel. %
1	9.82	9.35	95.18	9.63	98.11
2	8.65	8.96	103.55	8.16	94.27
3	5.64	5.18	91.86	5.57	98.80
4	9.61	9.66	100.58	10.38	108.01
5	12.36	12.31	99.62	11.86	95.94
6	9.02	9.01	99.81	9.57	105.99
7	4.63	4.95	106.93	4.62	99.82
8	20.09	21.40	106.54	19.60	97.54
9	9.18	9.65	105.05	10.01	109.00
10	7.44	7.89	106.10	8.20	110.26
Mean			101.77	-	101.52

Table 6. Main Pharmacokinetic Parameters of Test and Reference Methyldopa Drug.

Pharmacokinetic parameters	Test drug (Mean± SD)	Reference drug (Mean± SD)	Criteria of comparison	р
C _{max} , μg/ml	1.227±0.601	1.233±0.419	Student's t-test for dependent	0.952
$AUC_{0,t}, \mu g*h/ml$	6.219 ± 3.080	6.385 ± 2.153	group	0.740
AUC _{0-∞} , μg*h/ml	6.436 ± 3.181	6.529 ± 2.187		0.997
$C_{max}/AUC_{0-t}, h^{-1}$	0.202 ± 0.038	0.198 ± 0.037		0.663
K _{el} , h ⁻¹	0.229 ± 0.110	0.207 ± 0.107		0.127
T _s , h	3.89 ± 2.13	4.45±2.37		0.056
MRT, h	4.97 ± 0.77	4.98 ± 0.76		1.000
T_{max} , h	3.18 ± 1.23	2.90 ± 0.86	Wilcoxon signed-rank test	0.469

Note: C_{max} - maximum plasma concentration in blood; T_{max} - time-to-peak concentration; AUC_{0-t} - area under the pharmacokinetic concentration-time curve from time zero to the last blood sampling procedure; AUC_{0-x} - area under the pharmacokinetic curve from time zero to infinity; C_{max}/AUC_{0-t} - relative absorption rate; K_{cl} - terminal elimination rate constant; T_S - drug elimination half-life; MRT - mean residence time.




Figure 5. Averaged pharmacokinetic profiles of methyldopa concentrations in blood plasma after administration of test (T) and reference (R) drug.




Figure 6. Averaged pharmacokinetic profiles of MA and DMA concentrations in blood plasma after administering a single dose of Duspatalin (200 mg).

Table 7. Results of Comparison of Bioequivalence of Methyldopa formulations.

			90% confide	nce intervals		
	Geometric	Estimat	ed value	Accepta	ble value	Intra-subject
Parameter	mean ratio	Lower limit	Upper limit	Lower limit	Upper limit	variability coefficient
AUC _{0-t}	92.93%	80.69%	107.03%	80.00	125.00	29.08%
C_{max}	94.89%	80.88%	111.34%	80.00	125.00	33.10%
C_{max}/AUC_{0-t}	102.11%	93.95%	110.98%	80.00	125.00	16.92%

Table 8. Pharmacokinetic Parameters of Mebeverine Metabolites.

				Mebeve	erine acid				
	C _{max} , ng/ml	T _{max} , h	AUC _{0-t} , ng·h/ml	AUC _{0-∞,} ng·h/ml	$AUC_{0-t}/$ $AUC_{0-\infty}$, %	C _{max} / AUC _{0-t} , h ⁻¹	K _{el} , h-1	T _{1/2} , h	MRT, h
Mean±SD	62.52±	3.27±	293.94±	365.85±	82.34±	0.2238±	$0.27584 \pm$	2.87±	5.82±
	35.01	1.03	151.78	140.49	11.61	0.0598	0.08426	1.39	1.57
Min	22.3	1.5	54.03	110.01	47.43	0.1649	0.08683	1.61	4.27
Max	173.5	5	601.03	649.49	96.04	0.4128	0.43138	7.98	11.45
CV %	56.00	31.55	51.64	38.40	14.10	26.72	30.55	48.38	26.91
			D	esmethyl m	ebeverine aci	d			
			ATIC	ATTO	ATTO	<i>C</i> /			

	C _{max} , ng/ml	T _{max} , h	AUC _{0-t} , ng·h/ml	AUC _{0-∞,} ng•h/ml	AUC _{0-t} / AUC _{0-x} , %	C _{max} / AUC _{0.4} , h ⁻¹	$\mathbf{K}_{\mathrm{el}},\mathbf{h}^{\text{-}1}$	T _{1/2} , h	MRT, h
Mean±SD	$291.81 \pm$	$3.19\pm$	2191.85±	2551.74±	86.26±	$0.1339 \pm$	$0.11605 \pm$	$7.52\pm$	11.2±
	125.92	1.48	542.94	546.96	11.62	0.0497	0.06657	3.31	4.25
Min	114.6	1.5	1081.45	1388.65	40.34	0.0838	0.04631	2.33	5.34
Max	607.4	8	3262.43	3547.35	95.84	0.2969	0.29771	14.97	22.49
CV %	43.15	46.44	24.77	21.43	13.47	37.15	57.36	44.04	37.97

Conclusions

The method for the quantitative determination of substances forming such metabolites as glucuronides, N-oxides, esters, lactones and also for the simultaneous analysis of esters and lactones with their acid forms should ensure that there is no influence of degradation of these compounds in the ion source on the results of measuring the analyte concentration. The selection of storage conditions for samples containing substances with unstable functional groups should begin by selecting an anticoagulant, based on a study of short-term stability and freeze/thaw stability of the analyte and back-conversion of its meta-

bolites. When obtaining an unacceptable result, it is necessary to select a combination of an anticoagulant and a stabiliser solution, as well as a concentration of this solution and its volume ratio to the biological fluid. After this, the developed method should be validated using the selected anticoagulant or a combination of the anticoagulant and stabiliser solution.

Thus, the application of this approach to developing a bioanalytical method for determination of compounds containing unstable functional groups makes it possible to avoid obtaining false assay results. This significantly reduces the risk to the health of patients when administering generic drugs or when using the results of therapeutic drug monitoring for correcting the dosage.

References

- Benoit-Biancamano MO, Caron P, Levesque E, Delage R, Couture F, Guillemette C (2007) Sensitive High-performance Liquid Chromatography-tandem Mass spectrometry Method for Quantitative Analysis of Mycophenolic Acid and Its Glucuronide Metabolites in Human Plasma. Journal of Chromatography B 858: 159–167. https://doi.org/10.1016/j.jchromb.2007.08.023 [PubMed]
- Brandhorst G, Streit F, Goetze S, Oellerich M, Armstrong VW (2006) Quantification by Liquid Chromatography Tandem Mass Spectrometry of Mycophenolic Acid and Its Phenol and Acyl Glucuronide Metabolites. Clinical Chemistry 52(10): 1962–1964. https://doi.org/10.1373/clinchem.2006.074336 [PubMed]
- Brêtas JM, César IC, Brêtas CM, Teixeira LS, Bellorio KB, Mundim IM, Pianetti GA (2016) Development and Validation of an LC-ESI-MS/MS Method for the Simultaneous Quantification of Naproxen and Sumatriptan in Human Plasma: Application to a Pharmacokinetic Study. Analytical and Bioanalytical Chemistry 15: 3981–3992. https://doi.org/10.1007/s00216-016-9488-x [PubMed]
- Council of the Eurasian Economic Commission (2016) On Approval of the Rules for Conducting Bioequivalence Studies on Medicines in the Eurasian Economic Union. Decision of the Council of the Eurasian Economic Commission № 85 of November 3, 2016 № 85. http://docs.cntd.ru/document/456026107 [in Russian]
- Dell D (2004) Labile Metabolites. Chromatographia Supplement 59: 139–148. https://doi.org/10.1365/s10337-003-0169-5 [ResearchGate]
- Elliott S, Burgess V (2006) Investigative Implications of the Instability and Metabolism of Mebeverine. Journal of Analytical Toxicology 30: 91–97. https://doi.org/10.1093/jat/30.2.91 [PubMed]
- European Medicines Agency (2010) Committee for Medical Products of Human use: Guideline on Bioanalytical Method Validation.
 URL: www.ema.europa.eu/docs/en_GB/document_library/Scientific_guideline/2011/08/WC500109686.pdf [accessed: 5 February 2018)
- Heinig K, Bucheli F, Hartenbach R, Gajate-Perez A (2010) Determination of Mycophenolic Acid and Its Phenyl Glucuronide in Human Plasma, Ultrafiltrate, Blood, DBS and Dried Plasma Spots. Bioanalysis 8(2): 1423–1435. https://doi.org/10.4155/bio.10.99 [PubMed]
- Hilhorst M, Van Amsterdam P, Heinig K, Swanziger E, Abbott R (2015) Stabilization of Clinical Samples Collected for Quantitative Bioanalysis a Reflection from the European Bioanalysis Forum. Bioanalysis 7(3): 333–343. https://doi.org/10.4155/bio.14.290 [PubMed]
- Khokhlov AL (Ed.) (2017) Theoretical and Practical Basis for Conducting Studies of Generic Drugs. Moscow. Yaroslavl Prague: Photolife, 227 p. [in Russian]
- Khokhlov AL, Dzhurko YA., Yaichkov II, Shitov LN, Shitova AM. Miroshnikov AE, Khozova LA (2017a) The Rapid and Sensitive Hplc-Ms/Ms-Method of Determination of Mebeverine Metabolites in Human Plasma. Mathews Journal of Pharmaceutical Science 1(2): 010.
- Khokhlov AL, Dzhurko YA, Kubeљ V, Shitov LN, Yaichkov II, Shitova AM (2017b) Method of Quantitative Determination of Methyldopa in Human Plasma. Journal of Volgograd State Medical University 63(3): 105–108. https://doi.org/10.19163/1994-9480-2017-3(63)-105-108 [in Russian]
- Khokhlov AL, Yaichkov II, Dzhurko YA, Shitov LN (2017c) Methodical Approaches to Bioassay of Phenolic Hydroxylenes Contain Substances. Medical News of North Caucasus 12(3): 294–299. https://doi.org/10.14300/mnnc.2017.12088

- Khokhlov AL, Yaichkov II, Shitova MA, Shitov LN, Dzhurko YA, Miroshnikov AE (2016) A Comparative Study of Pharmacokinetics of Coated Tablets of Mycophenolic Acid. Pharmacokinetics. Pharmacodynamics 4: 3–8 [in Russian]
- Kristinsson J, Snorradbttir I, Johannsson M (1994) The Metabolism of Mebeverine in Man: Identification of Urinary Metabolites by Gas Chromatography/Mass Spectrometry. Pharmacology and Toxicology 14: 174–180. [PubMed]
- Li W, Zhang J, Tse F (2011) Strategies in Quantitative LC-MS/MS Analysis of Unstable Small Molecules in Biological Matrices. Biomedical Chromatography 25: 258–277. https://doi.org/10.1002/bmc.1572 [PubMed]
- Li W, Zhang J, Tse F (2013) Handbook of LC-MS Bioanalysis. New Jersey: Jonh Wiley and Sons, 675 p.
- Liu Q, Jiao Z, Zhong M, Zhang M, Geng F, Zhao H (2017) Effect of Long-term co-administration of Compound Glycyrrhizin Tablets on the Pharmacokinetics of Mycophenolic Acid in Rats. Xenobiotica 46(7): 627–633. https://doi.org/10.3109/00498254.2015.1103386 [PubMed]
- Ministry of Healthcare of the Russian Federation (2016) On Approval of the Rules for Good Laboratory Practice. Order of the Ministry of Healthcare of the Russian Federation of 01.04.2016 №199н. http://cdnimg.rg.ru/pril/130/25/64/43232.pdf [in Russian]
- Mironov AN (2014) Guidance on Inspection of Medicines. Moscow: Polygraph Plus 1: 328 p. [in Russian]
- Oliveira CH, Barrientos-Astigarraga RE, Sucupira M, Graudens GS, Muscaró MN, Nuccia G (2002) Quantification of Methyldopa in Human Plasma by High-performance Liquid Chromatography-electrospray Tandem Mass Spectrometry. Application to a Bioequivalence Study. Journal of Chromatography B 768(2): 341–348. https://doi.org/10.1016/S1570-0232(01)00612-2 [ResearchGate]
- Patel DS, Sharma N, Patel MC, Patel BN, Shrivastav PS, Sanyal M. Sensitive and Selective Determination of Diffunisal in Human Plasma by LC –MS. Journal of Chromatographic Science. 2013;51:872–882. https://doi.org/10.1093/chromsci/bms181 [PubMed]
- Róna K, Ary K, Renczes G, Gachályi B, Grézal GY, Drabant S, Klebovich I (2001) Comparative Bioavailability of Alpha-Methyldopa Normal and Film Tablet Formulations after Single Oral Administration in Healthy Volunteers. European Journal of Drug Metabolism and Pharmacokinetics 26(1–2): 25–30. https://doi.org/10.1007/BF03190372 [Full text]
- Sengupta P, Chatterjeea B, Mandala UK, Gorain B, Pal TK (2017) A High Throughput LC-MS/MS Method for Simultaneous Quantitation of Pioglitazone and Telmisartan in Rat Plasma; Development, Validation and Pharmacokinetic Application. Journal of Pharmaceutical Analysis7(6): 381–387. https://doi.org/10.1016/j.jpha.2017.05.004 [ResearchGate]
- Shah I, Barker J, Naughton DP, Barton SJ, Ashraf SS. Determination of Diclofenac Concentrations in Human Plasma Using a Sensitive Gas Chromatography Mass Spectrometry Method. Chemistry Central Journal. 2016;10:1-10. https://doi.org/10.1186/s13065-016-0199-3 [PubMed]
- Valizadeh H, Nemati M, Hallaj-Nezhadi S, Ansarin M, Zakeri-Milani P (2010) Single Dose Bioequivalence Study of A-methyldopa Tablet Formulations Using a Modified HPLC Method. Arzneimittelforschung 60(10): 607–611. doi: 10.1055/s-0031-1296333 [PubMed]

- Veeragoni AK, Sindgi VM, Satla SR (2016) Bioanalytical Validated LC-MS Method for Determination of Naproxen in Human Plasma.
 International Journal for Modern Trends in Science and Technology 2(19): 96–99. [Full text]
- Vlase L, Mihu D, Popa D-S, Popa A, Briciu C, Loghin F, Ciortea R, Mihu C (2013) Determination of Methyldopa in Human Plasma by LC/MS-MS for Therapeutic Drug Monitoring. Studia Universitatis Babes-Bolyai, Chemia 58(1): 31–41. [ResearchGate]
- Yaichkov II, Khokhlov AL, Dzhurko YA, Shitov LN, Trubnikov AA (2017) Methods of Stabilizing Drug Substances and Their Metabolites in Biological Fluids in Bioanalytical Studies (Review). Development and Registration of Medicines 2: 160–164. [in Russian]
- Ye X, Bishop AM, Reidy JA, Needham LL, Calafat AM (2007) Temporal Stability of the Conjugated Species of Bisphenol A, Parabens, and Other Environmental Phenols in Human Urine. Journal of Exposure Science and Environmental Epidemiology 17: 567–572. https://doi.org/10.1038/sj.jes.7500566 [PubMed]

Author Contributions

- Alexander L. Khokhlov, Doctor of Medical Sciences, Professor, Corresponding member of The Russian Academy of Sciences, Head of the Department of Clinical Pharmacology, e-mail: al460935@yandex.ru. The author set the goals and objectives of the clinical part of the bioequivalence and pharmacokinetics studies and developed the design of these studies.
- Ilya I. Yaichkov, post-graduate student of the Department of Clinical Pharmacology, Junior Research Fellow of M.V. Dorogov Centre for Transfer of Pharmaceutical Technologies, Yaroslavl State Pedagogical University named after K.D. Ushinsky, e-mail: ilya_1993_08@mail.ru. The author set the goal and objectives of the bioanalytical part of the study, did a literature review, took part in developing and validating the methods, as well as in carrying out the bioassay using the methods.
- Yuriy A. Dzhurko, PhD in Pharmacy, Senior Analyst, Quinta-Analytica Yaroslavl LLC, e-mail: y.dzhurko@qayar. ru. The author took part in developing and validating the methods, as well as in carrying out the bioassay using the methods.
- Leonid N. Shitov, PhD in Biology, Head of Bioanalytical laboratory Quinta-Analytica Yaroslavl LLC, Assistant at the Department of Polyclinical Therapy and Clinical Laboratory Diagnostics, Yaroslavl State Medical University, e-mail: schitov@inbox.ru. The author took part in developing and validating the methods, as well as in carrying out the bioassay using the methods.
- Anastasia M. Shitova, PhD in Physics and Mathematics, statistician Quinta-Analytica Yaroslavl LLC, e-mail: am_schitova@cphinvest.ru. The author did the statistical processing of the data obtained in the bioequivalence and pharmacokinetics studies.